

InControl 1050

Quality – Made in Germany



Independent measuring device for control and evaluation of CO₂, temperature and O₂

Features:

- Accurate measuring of CO₂, temperature and optional O₂
- Single or continuous measurement in arbitrary intervals
- Measurement and documentation of multiple incubators
- Documentation of measurements with date, time and incubator number
- Easy menu navigation
- Continuous operation by Li-Ion battery or mains adapter possible
- Data-Download via USB; PC software DataVISUAL`09 incl.
- Compatible to Windows XP and Windows 7

Scope of supply:

- InControl 1050 including Transport case
- Mains adapter
- Tubing, adapter and condensation trap
- PC software: DataVISUAL`09 (on USB-Stick)
- USB data cable

REF 14709



Technical data

Case:

- Plastic (ABS)

CO₂ measurement:

- Dual beam IR sensor
- Range: 0 – 10 % CO₂
- Display definition: 0.1 % CO₂
- Accuracy:
 - 0 – 6 % CO₂ ± 0.2 % CO₂
 - 6 – 10 % CO₂ ± 0.3 % CO₂
- Sample volume per measurement (2 min.): approx. 0.8 l

Temperatur measurement:

- Platinum-temperature sensor PT1000
- Range: 0 – 100 °C
- Display definition: 0.1 °C
- Accuracy:
 - 20 – 50 °C ± 0.2 °C
 - 50 – 100 °C ± 0.3 °C

O₂ measurement:

- Galvanic sensor
- Range: 0 – 100 % O₂
- Display definition: 0.1 % O₂
- O₂-sensor not included

Data logger:

- Recording of measurement logging in arbitrary intervals
- Output of measurements via:
 - backlit LCD
 - PC software Data Visual'09 (via USB data cable)
- Interval for data logging: 15 - 120 min. adjustable in steps of 5 minutes
- storage of max. 1008 measurements with date and time

Power supply:

- Rechargeable Li-Ion battery
- Power supply 100 – 240 V AC, 5 V DC 1.5 A

Operation time:

- Use of Rechargeable battery approx. 4 h
- Time to recharge battery approx. 4 h



Optional accessories:



Sensor for O₂ measurement

REF 10556



Flat tape sensor PT1000 for temperature measurement inside Labo C-Top or other benchtop incubators

REF 15064



Surface sensor for temperature measurement

REF 10537



Rev. 6_08/2017